

# CERTIFICATE OF CALIBRATION


This Certificate certifies that the items listed below have been tested, manufactured, calibrated and/or inspected in accordance with IDEAL INDUSTRIES procedures. These procedures are ISO 9001:2008 controlled and are designed to assure that the instrument will meet its published specification.

Inspections and/or tests and Calibration have been performed, as applicable at or on behalf of

### IDEAL INDUSTRIES PRODUCT SOLUTIONS

Stokenchurch House, Oxford Road  
Stokenchurch, High Wycombe  
Buckinghamshire. HP14 3SX. UK

Calibration compliance is achieved by comparison to a calibrated standard, using an Automated Test Equipment (ATE) Station, in accordance with IDEAL LanTEK® System Factory Calibration Procedure 160626/160609 on file at the above address. Procedure 160637 details the method of calibration showing traceability of calibration of the ATE with a Network Analyser with calibration certificates documenting traceability to the National Physical Laboratory (NPL) and or National Institute of Standards and Technology (NIST).

Reference No:	LRIE-AHWCJ7
Customer Name:	INTELTEC
Model No:	LanTEK II-1000
Display Handset Serial Number	1036182
Remote Handset Serial Number	1036232
Firmware Version	2.050
Equipment used for Calibration:	LanTEK ATE Calibration station (Serial No: 14)
Date of Calibration:	31 January 2017
Recommended Next Calibration:	03 March 2018
Environmental Conditions	Temperature: 22°C ± 2°C
No of Pages	2
Calibrated by:	Thurai Siva - IDEAL UK
Signature:	
Date:	31/01/2017